



America

CERTIFICATE


No. U8 11 04 76134 003

Holder of Certificate: Quantum Dental Technologies Inc.

748 Briar Hill Ave
Toronto ON M6B 1L3
CANADA

Production Facility(ies): 73765, 66506

Certification Mark:



C NRTL US

Product: General Medical Devices
Dental laser system

Model(s): The Canary System

Parameters:

Rated Input Voltage:	100-120 V/200-240 V
Rated Frequency:	50/60 Hz
Rated Input Power:	100 VA
Protection Class:	I
Degree of Protection:	Type B

Tested according to:

CAN/CSA C22.2 No. 601.1-M90 excluding:
EMC (Clause 36), biocompatibility (clause 48)
programmable electronic systems (clause 52.1
as it pertains to IEC 60601-1-4).
CAN/CSA C22.2 No. 60601-2-22-01 (R2005)
UL 60601-1:2003 excluding:
EMC (Clause 36), biocompatibility (clause 48),
programmable electronic systems (clause 52.1
as it pertains to IEC 60601-1-4).
IEC 60601-2-22:1995

The product was voluntarily tested according to the relevant safety requirements and mentioned properties. It can be marked with the certification mark shown above. The certification mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles that described by ISO/IEC Guide 67, Conformity assessment - Fundamentals of product certification, System 3. See also notes overleaf.

Test report no.: KI1100346-000

Date, 2011-04-25

Page 1 of 2

